

# BLOMDAHL

FEELGOOD JEWELLERY

## **Recall of Medical Plastic Daisy Crystal (art. no 12-0114-41) for Ear piercing LOT 002ZP**

There is a fault in the specific gluing batch regarding the ear piercing product **12-0114-41 with LOT 002ZP**, which therefore makes the stones fall out of its place because the glue has not hardened. This means that there will be a risk that the glue will come into contact with the skin.



The recall only applies to Medical Plastic Daisy in the colour crystal/white, with **LOT 002ZP and ear piercings that have been made after the 18 th of May 2022**. Please contact the store who performed the ear piercing if you believe that you are affected by this recall, or if you can see that stones have fallen out. Remove and discard the earrings and contact the store that made the ear piercing. Then you will be offered a new ear piercing free of charge as soon as the old ear piercing has healed. You will also be offered a free pair of skin friendly jewellery in price category A.

Please contact us if you have any questions - [customerservice@blomdahl.com](mailto:customerservice@blomdahl.com)

We are very sorry for this and the inconvenience this causes and thank you for your help in this matter.